

Manufacturer's Certificate of Validation

Certificat No.: MCV110512001

Customer: Company
Address
Aera Code and City
Country

imec Messtechnik GmbH certifies that the products mentioned on the following page(s) have been thoroughly tested and validated in accordance with China Verification Regulation of Recorders for Industrial-Process (JJG 74-2005) and have met performance accuracy specifications over the stated ranges.

Reference Instrumentation

| Device | Model | Accuracy | Serial Number |
|---------------------------------|-------------|---------------------|---------------|
| Thermometer readout | Fluke 1502A | + -0.006@0°C | A23609 |
| Refrigerated/Heating Circulator | XORTS-40A | 0.01°C (volatility) | 11-061 |
| SPRT | AWT1842 | Class 1 | 00096 |

Reference instrumentation used for validation is traceable to PMTI, a CNAS accredited laboratory (No.L0786), according to Mutual Recognition Agreement (MRA), they are also traceable to NIST also.

Environment Conditions

| Air Temperature | Relative Humidity | Pressure |
|-----------------|-------------------|------------|
| 22.1°C | 54% | 998+-5mbar |

Products

| Product Name | Model | Serial Number |
|----------------------------------|---------|--|
| ITAG-SingleUse Temperature Label | GS60DCN | 20314314-20314320, 20314301-20314303 (10 units) |

Validation approved by:
Date: May 24th, 2011



Stamp:



Messtechnik GmbH
Lilienthalstrasse 23
74078 Heilbronn
Tel. 0 70 66 - 9 15 02 30
Fax 0 70 66 - 9 15 02 33

*this certificate is valid for a period of 1 (one) year from date of sale.